



Australian Government
Department of Health
Therapeutic Goods Administration

Certificate of Official Analyst

I, Elisabeth Kerr, being an Official Analyst appointed under Regulation 25 of the Therapeutic Goods Regulations 1990, certify that a sample of therapeutic goods, bearing TGA sample number 1602000675, has been examined and analysed.

ARTG Number: 156690
Product : HeartSine Samaritan PAD SAM 500P Automated External Defibrillator
Sponsor : Aero Healthcare
Batch Number : SN16B00511031 Expiry Date: N/A

Test (Standard Subclause)	Result	Requirement
Automatic External Defibrillators (AED) testing to standard IEC 60601-2-4		
Energy Accuracy (201.12.1) - Adult Mode	Tested	
Energy Accuracy (201.12.1) - Paediatric Mode	Tested	
Output Voltage (201.12.4.101)	Passed	<5000V
Charge Time (201.101.4) - At 7th Discharge	Pass	<35s
Charge Time (201.101.4) - At 16th Discharge	Pass	<40s
Rhythm Recognition (201.107)	Tested	

Certificate : 1602000675

Official Analyst Signature :

Page 1 of 1

Issued 28-September-2016